

MINUTES OF PRE-BID MEETING OF ANNUAL TENDER OF MAYO HOSPITAL, LAHORE FOR F.Y 2026-27 OF SURGICAL DISPOSABLE/MEDICAL DEVICES/LAB KITS REAGENTS & EQUIPMENT PLACEMENT BASIS / RADIOLOGY/ UROLOGY/ EYE DISPOSABLE ITEMS ETC, HELD ON 01.06.2026 HELD IN THE CONFERENCE ROOM, SURGICAL TOWER AT 10:00AM.


The meeting was commenced at the scheduled date and time in conference room Surgical Tower Mayo Hospital Lahore. The purpose of the meeting was to ensure the widest competition on compliance with the Punjab Procurement Rule 2014 and addressed the observations from the pre-bid discussion under EPADS.


The following amendments were unanimously recommended by the end user, endorsed by the committee, and approved by the Competent Authority. These amendments will be considered as a part of bidding document for the F.Y 2026-27.


FIRM NAME	Advertised Specifications	Amended Specifications
Lab Kits Reagents & Equipments Placement Basis (A05-B)	Compliant with a semi-automated machine	To Uphold the advertised specification in bidding document
	Inclusion of seasonal/low running parameters such as H.Pylori and Dengue within the comprehensive immunoassay list	To Uphold the advertised specification in bidding document
	Specifications limited exclusively to High-Performance Liquid Chromatography (HPLC.)	To Uphold the advertised specification in bidding document
Lab Kits Reagents & Equipments Placement Basis (A05-B)	More than 30 cooled reagents position on Board	More than 20 cooled reagents position on Board
Lab Kits Reagents & Equipments Placement Basis (A05-B)	Instrument and reagents both will be CE Marked/ FDA Approved	The Bidder shall provide reagents, calibrators, controls, and consumables of the same manufacturer/ brand as the offered on all instruments advertised in the Tender.
Lab Kits Reagents & Equipments Placement Basis (A05-B)	instrument should be brand new (not refurbished) and CE marked/ FDA approved	To Uphold the advertised specification in bidding document
Lab Kits Reagents & Equipments Placement Basis (A05-B)	Throughput: 170 Tests /hr or above	To Uphold the advertised specification in bidding document
	Throughput: 50 tests/hr or more	To Uphold the advertised specification in bidding document
	HCV PCR = Rs. 1350/ test HBV PCR = Rs. 1400/ test	To Uphold the advertised estimated cost in bidding document.
	Restrictive Requirement of "US FDA Approved / CE Marked/ WHO Prequalified"	To Uphold the advertised specification in bidding document


	The system should have minimum capacity of 32 samples with maximum run time 30 mins or less	The system should have minimum capacity of 32 samples with maximum run time 60 mins or less
Tender: General Surgical Disposable items (A14)	Diosol Hydrogen Peroxide 10% or above with silver ions disinfectant Fumigation (5 Liter)	To Uphold the advertised specification in bidding document
Tender. Urology Disposable items (A14)	Dj stent 4.7 FR with guide wire 0.032 FDA Approved (Adult) 4.7FR	Dj stent 4.7/4.8 FR with compatible guide wire 0.032 CE/FDA Approved.
	Ureteric Catheter Open End (Adult) 6 FR with stylet	Ureteric Catheter Open End Adult 6 FR with/without Stylet CE/FDA Approved
	Ureteric Catheter Open End Paeds 4.7 FR with stylet	Ureteric Catheter Open End Paeds 4.7FR/5FR with/without Stylet CE/FDA Approved


Handwritten marks: a checkmark and the number '2'.



Mr. Muhammad Hifzan
 Audit Officer, Mayo Hospital Lahore



Ms. Kanwal Javed
 Deputy Drug Controller
 Mayo Hospital Lahore


Mst. Anila Saeed
 Deputy Drug Controller
 Mayo Hospital Lahore


Dr. Rabia Rathore
 Associate Professor of Medicine/Head of (WMW)
 KEMU/ Mayo Hospital Lahore


Dr. Qazi Mumtaz Ahmad
 Assistant Professor of Pediatric Medicine (Unit-I)
 KEMU/ Mayo Hospital Lahore


Prof. Dr. Nasir Ahmad Chahudary
 Professor of Ophthalmology /HOD (Unit-II)
 KEMU/ Mayo Hospital Lahore Chairman


 All attached documents
 Checked & Verified by above
 members as per rules.
 uplead please
Chief Executive Officer
Mayo Hospital, Lahore.
 11 JUN 2025